

Virtual Site Visit - Site Health Report










Site Name: NORTHSIDE
 Address: 565 Westminster Pike

Starting Date: 02/01/2011
 Ending Date: 02/07/2011

Primary Measurements

<u>Parameter</u>	<u>Min</u>	<u>Max</u>	<u>Average</u>
SO2_PPM	1.7	3.2	1.9
NOY	3.2	11.6	7.7
NOY_DIF	0.23	0.54	0.31
CO_TRACE	0.04	1.12	0.76

Secondary Measurements

<u>Parameter</u>	<u>Min</u>	<u>Max</u>	<u>Average</u>	<u>Range</u>
ShelterTemperature	21.2	24.3	22.7	
SO2_PPM:SampleFlow	0.62	0.68	0.67	
SO2_PPM:SamplePress	723	776	757	
SO2_PPM:ChamberTemp	24.6	29.3	28.5	
NOY: SampleFlow	0.41	0.66	0.65	
NOY: SamplePress	717	782	755	
CO_TRACE: SampleFlow	0.60	0.67	0.65	
CO_TRACE: SamplePress	719	781	759	
CO_TRACE: BoxTemperature	24.3	36.3	28.2	

Site Logbook

<u>Category</u>	<u>User Name</u>	<u>Entry Time</u>	<u>Entry</u>
Adjustment	AirVision	2/23/2011 9:43:37 AM	CO analyzer zero offset adjusted + 0.3